

Detection of c.586C>T mutation in exon 2 of
SLC3A1 gene causing cystinuria in
Newfoundland and Landseer breeds

Customer: Jan Novák, Dlouhá 1, 30000 Plzeň, Czech Republic

Sample:

Sample: 21-12345

Date received: 01.02.2021

Sample type: blood

Information provided by the customer

Name: Lassie DEMO

Breed: Plemeno

Tattoo number: 1392013

Microchip: 123 456 789 012 345

Reg. number: REGQ12345

Date of birth: 1.1.2020

Sex: female

Date of sampling: 01.02.2021

The identity of the animal has been checked.

Result: Mutation was not detected (N/N)

Legend: N/N = wild-type genotype. N/P = carrier of the mutation. P/P = mutated genotype (individual will be most probably affected with the disease). (N = negative, P = positive)

Explanation

Presence or absence of c.586C>T mutation in exon 2 of SLC3A1 gene causing cystinuria in Newfoundland and Landseer breeds was tested. Due to high concentration of cystine, hard crystals of cystine can be formed in the urine of affected individuals. The crystals can cause uncomfortable health complications e.g. urinary deposits and/or kidney stones, which can ultimately lead to the obstruction of the urinary tract.

Mutation that causes cystinuria is inherited as an autosomal recessive trait. That means the disease affects dogs with P/P genotype only. The dogs with N/P genotype are considered carriers of the disease (heterozygotes). In offspring of two heterozygous animals following genotype distribution can be expected: 25 % N/N, 25 % P/P and 50 % N/P.

Method: SOPAgriseq_canine, ngs, accredited method

Sensitivity (probability of correct identification of the defective form of the gene in heterozygous or mutated homozygous) is higher than 99%. Specificity (probability of correct identification of the normal form of the gene in a normal homozygous or heterozygous) is higher than 99%.

Date of issue: 06.02.2021

Date of testing: 01.02.2021 - 06.02.2021

Approved by: Mgr. Martina Šafrová, Laboratory Manager



Genomia is accredited in compliance with ISO/IEC 17025:2018 under #1549

Genomia s.r.o, Republikánská 6, 31200 Plzeň, Czech Republic

www.genomia.cz, laborator@genomia.cz, tel: +420 373 749 999



Report verification code is: 12AB-CD34-GENO-MIA0-EFGH. You can verify report online at www.genomia.cz

Without a written consent by the lab, the report must not be reproduced unless as a whole.

The result refers only to the tested sample, as received. Genomia is not responsible for the accuracy of the information provided by the customer.